



* Please read the instructions carefully before use

INTENDED USE

Artron COVID-19 Antigen Test is a rapid and convenient lateral flow immunochromatographic assay for the qualitative detection of SARS-CoV-2 nucleocapsid protein from nasopharyngeal or nasal swab samples obtained from individuals suspected of COVID-19 by their healthcare provider within five to seven days of symptom onset or individuals without symptoms or other epidemiological reasons to suspect COVID-19 infection, when tested twice over 2 (or 3) days with at least 24 hours (and no more than 36 hours) between tests. The rapid test device is for professional and point of care use only and is intended to be used as an aid in the diagnosis of SARS-CoV-2 infection.

This assay provides preliminary test results. Positive results indicate the presence of viral antigens, but the clinical correlation with patient history and other diagnostic information is necessary to determine infection status. Positive results do not rule out a bacterial infection or co-infection with other viruses. The agent detected may not be the definite cause of disease. Negative results from asymptomatic patients suspected of SARS-CoV-2 exposure and patients with symptom onset beyond seven days, should be treated as presumptive and confirmation with a molecular assay, if necessary, for patient management, may be performed. Negative results do not rule out SARS-CoV-2 infection and should not be used as the sole basis for treatment or patient management decisions, including infection control decisions. Negative results should be considered in the context of a patient's recent exposures, history and the presence of clinical signs and symptoms consistent with COVID-19.

Artron COVID-19 Antigen Test may be used in any laboratory and non-laboratory environment that meets the requirements specified in the instructions for use and local regulation. This product is intended for use by healthcare professionals in clinical laboratories or Point of Care (POC) settings. The result of this test should not be the sole basis for the diagnosis and the test results should be confirmed by local government approved Real-Time Reverse Transcriptase (RT)-PCR Diagnostic kit.

SUMMARY AND PRINCIPLE OF THE ASSAY

Severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2) is the virus strain that caused an outbreak of a novel coronavirus disease (COVID-19), which has subsequently affected countries and regions worldwide. Severe disease onset might result in death due to massive alveolar damage and progressive respiratory failure. On March 11, 2020, the World Health Organization (WHO) has declared the global outbreak of COVID-19 a pandemic associated with substantial morbidity and mortality.

Artron COVID-19 Antigen Test is an antigen-capture immunochromatographic assay, detecting presence of SARS-CoV-2 nucleocapsid protein in nasopharyngeal swab specimens. SARS-CoV-2 specific antibody and a control antibody are immobilized onto a membrane support as two distinct lines-Test line(T) and Control line(C) and combined with colloidal gold- monoclonal antibody against SARS-CoV-2 antigen deposited on the conjugate pad to construct a test strip. When the swab sample migrates in the test strip, SARS-CoV-2 nucleocapsid protein bind to anti-SARS-CoV-2 nucleocapsid protein antibody-gold conjugate, forming an immune complex. The immune complex is then captured by the test line on the nitrocellulose membrane as it migrates through the strip, forming a visible pink or purple line, indicating positive result. If SARS-CoV-2 are absent in the sample, no pink or purple line will appear in the test line, indicating a negative result.

To serve as an internal process control, a control band was designed to indicate that the test is performed properly. This control line should always be seen after test is completed. Absence of a control line in the control region is an indication of an invalid result.

PACKAGE CONTENTS

- Test cassettes with desiccant in individual pouch (Catalog No. A03-50-422P): 1 device for 1pc/pack, 5 devices for 5 pcs/pack, 25 devices for 25 pcs/pack.
- Extraction tubes sealed with sample extraction buffer (300µL/tube); 1 tube for 1pc/pack, 5 tubes for 5 pcs/pack, 25 tubes for 25 pcs/pack.

- Extraction tube caps; 1 cap for 1pc/pack, 5 caps for 5 pcs/pack, 25 caps for 25 pcs/pack.
- Sterilized nasopharyngeal swabs (Catalog No. 96000) for Catalog A03-50-422PNP1, A03-50-422PNP5 and A03-50-422PNP25 for nasopharyngeal swab specimens; 1 nasopharyngeal swab for 1pc/pack, 5 nasopharyngeal swabs for 5 pcs/pack, 25 nasopharyngeal swabs for 25 pcs/pack.
- Sterilized nasal swabs (Catalog No. CF 075-P 3 B) for Catalog A03-50-422PNS1, A03-50-422PNS5 and A03-50-422PNS25 for nasal swab specimens; 1 nasal swab for 1pc/pack, 5 nasal swabs for 5 pcs/pack, 25 nasal swabs for 25 pcs/pack.
- 1 tube rack for 25 pcs/pack.
- 1 Instruction for Use

MATERIALS REQUIRED (BUT NOT PROVIDED)

- Personal protective equipment
- Timer
- Biohazard container

WARNINGS AND PRECAUTIONS

- For in vitro diagnostic use only.
- The test is designed only for the detection of nasopharyngeal swab and nasal swab specimens.
- This test is only for the detection of proteins from SARS-CoV-2, not for any other viruses or pathogens.
- Do not reuse.
- Do not use if the pouch seal or its packaging is compromised.
- Do not use after the expiration date shown on the pouch.
- Do not mix and interchange different specimens.
- The swabs in the kits are approved for use with Artron COVID-19 Antigen Test. Do not use other swabs.
- If the test is stored refrigerated, ensure that the test units are brought to room temperature (15-30°C) at least 30mins before performing testing.
- Immediately use after opening the test device in the pouch.
- Complete the test within 1 hour after the reagent is opened.
- In order to obtain accurate results, the test must follow this package insert.
- Wear personal protective equipment such as laboratory coats, disposable gloves and eye protection when running each test and handling patient specimens. Change gloves between handling of specimens suspected of COVID-19.
- Wash hands thoroughly after finishing the tests.
- Do not eat, drink, or smoke in the area where the specimens or kits are being handled.
- Clean up spills thoroughly with appropriate disinfectants.
- Handle all specimens as if they contain infectious agents. Observe established precautions against microbiological hazards throughout testing procedures.
- Dispose of all specimens and used devices in a proper bio-hazard container. The handling and disposal of the hazardous materials should follow local, national, or regional regulations.
- Keep out of children's reach.
- If the extraction buffer contacts the skin or eye, flush with copious amounts of water.

SPECIMEN COLLECTION AND PREPARATION





Note:

Before proceeding with sample collection and testing, please read the instruction carefully, and operate strictly in accordance with the instructions.


Freshly collected specimens should be processed immediately. Specimens in Artron sample extraction buffer are stable for up to 4 hours at 2-8°C or room temperature.

1. Tear off the aluminum foil seal from the extraction tube.




<p>2. Before collecting the sample, place the sample extraction tube into the tube holder on the box (for 1pc/pack and 5 pcs/pack) or provided tube rack</p>			
<p>3. Remove a swab from the pouch.</p>			
<p>4a. For Nasopharyngeal Swab Specimen collection</p>			
 <p>1) Insert a sterile swab into the nostril of the patient, reaching the surface of the posterior nasopharynx.</p>	 <p>2) Swab over the surface of the posterior nasopharynx.</p>	 <p>3) Withdraw the sterile swab from the nasal cavity.</p>	
<p>4b. For Nasal Swab Specimen collection</p>			
 <p>1) Insert the entire absorbent tip of the swab into your nostril, but do not insert the swab more than 3/4 of an inch (1.5 cm) into your nose.</p>	 <p>2) Slowly rotate the swab in a circular path against the inside of your nostril at least 5 times for a total of 15 seconds. Be sure to collect any nasal drainage that may be present on the swab.</p>	 <p>3) Gently remove the swab.</p>	 <p>4) Using the same swab, repeat steps 1 - 3 in your other nostril.</p>
<p>5. Insert the swab in the extraction tube. Swirl the swab tip vigorously in the buffer fluid at least 10 times.</p>			
<p>6. Remove the swab by rotating against the extraction tube while squeezing the sides of the tube to release the liquid from the swab. Properly discard the swab.</p>			
<p>7. Close the extraction tube with the provided extraction tube cap and push firmly onto the tube.</p>			

TEST PROCEDURES

<p>1. Get the test cassette from the sealed pouch by tearing at the notch and place the cassette on a flat, dry surface</p>	
<p>2. Hold the extraction tube vertically above the sample well, slowly add 4 drops of the specimen without air bubbles into the sample well. DO NOT touch the card with the dropper tip while dispensing</p>	
<p>3. Read and interpret the test result within 15-30 minutes. The test result should not be read and interpreted after 30 minutes. <u>If a test shows a negative result at the 15 minutes, not to discard the device immediately as some positive results may develop later in the 15-30 min interval.</u></p>	 <p>15-30 Min</p>  <p>DO NOT INTERPRET RESULTS AFTER 30 MINUTES.</p>
<p>4. All used test components should be disposed of in Biohazard Container.</p>	

RESULT INTERPRETATION



Negative:

A clear pink or purple colored band appears only at the control region (C), indicating a negative result.

Positive:

A clear pink or purple control band (C) and a detectable test band (T) appears, indicating a positive result.

Invalid:

No visible band appears at the control region. Repeat with a new test kit. If the test still fails, please contact the distributor with the lot number.

QUALITY CONTROL

Although the testing device contains an internal quality control (pink or purple colored band in the control region), good laboratory practice recommends the daily use of an outside control to ensure proper testing device performance. Quality control samples should be tested according to the standard quality control requirements established by your laboratory.

STORAGE AND STABILITY

- Test device in the sealed pouch can be stored at 2-30°C up to the expiration date. Do not freeze the test device.
- The test device should be kept away from direct sunlight, moisture, and heat.
- Shelf life: 18 months.

LIMITATIONS

- The test is only intended for nasopharyngeal swab and nasal swab specimens that are collected and tested directly, not for swab specimens stored in virus transport media.
- Failure to follow the Test procedures may adversely affect test performance and/or invalidate the test result.
- Inadequate or inappropriate sample collection, storage, and transport may yield false test results.

- A negative test result may occur if the level of antigen in a sample is below the detection limit of the test.
- False results may occur if specimens are tested past 4 hours of collection. Specimens should be tested as quickly as possible after specimen collection.
- The freshly collected specimens can be stably stored in the sample extraction buffer at room temperature up to 4 hours of collection.
- False negative results may occur if inadequate extraction buffer is used (e.g., <300µl) or inadequate specimen is added in the sample well (e.g., <4 drops).
- False negative results may occur if specimen swabs are not twirled sufficiently in the sample extract buffer.
- Positive test results do not rule out co-infections with other pathogens.
- Negative test results do not rule-out possible other non-COVID-19 viral infections.
- Negative results, from asymptomatic patients suspected of SARS-CoV-2 exposure and patients with symptom onset beyond seven days, should be treated as presumptive and confirmation with a molecular assay, if necessary, for patient management, may be performed.
- Results from antigen testing should not be used as the sole basis to diagnose or exclude SARS-CoV-2 infection or to determine infection status.
- Results from the test should be correlated with the clinical history, epidemiological data and other data available to the clinician evaluating the patient.
- This test only provides qualitative test result and cannot provide information about the virus concentration in the sample.
- The performance of the device has not been assessed on specimens from individuals who have been infected with emerging variants of SARS-CoV-2 of public health concern.
- For mutant virus strains or virus strains from different regions, the detection ability of the device may be different, which may lead to false negative.
- If the differentiation of specific SARS viruses and strains is needed, additional testing, in consultation with state or local public health departments, is required.
- The performance of this device has not been assessed in a population vaccinated against COVID-19.
- Positive and negative predictive values are highly dependent on prevalence. False negative test results are more likely during peak activity when prevalence of disease is high. False positive test results are more likely during periods of low SARS-CoV-2 activity when prevalence is moderate to low.
- Clinical studies in asymptomatic patients using serial testing are ongoing to establish clinical performance.
- The performance of this test has not yet been clinically validated for use in patients without signs and symptoms of respiratory infection or for serial screening applications. Note that performance may differ in these populations.

PERFORMANCE CHARACTERISTICS

• Limit of Detection (LoD)-Analytical Sensitivity

The limit of detection (LoD) of Artron COVID-19 Antigen Test is 1×10^3 TCID₅₀/mL.

• Cross Reactivity

None of the below related pathogens: Coronavirus OC43(ATCC: VR-1558™); Coronavirus NL63; Coronavirus 229E; SARS Coronavirus (2003-00592 strain); MERS Coronavirus(Florida/USA-2_Saudi Arabia_2014); H1N1 influenza virus (2009) (Canada/629/09 strain); H1N1 influenza virus (ATCC:VR-98™);Seasonal H3N2 influenza virus (Brisbane/10/07 strain); Influenza B (Yamagata/16/88 strain); Influenza B (Victoria/2/87 strain); Parainfluenza virus type 1(ATCC: VR-94™); Parainfluenza virus type 2 (ATCC: VR-92™); Parainfluenza virus type 3(ATCC: VR-93™); Parainfluenza virus type 4b (ATCC: VR-1377); Respiratory syncytial virus (ATCC: VR-1580™); Rhinovirus A (73)(ATCC: VR-1183™); Rhinovirus B (B42); Adenovirus type 1 (C); Adenovirus type 2 (C); Adenovirus type 3 (B); Adenovirus type 4; Adenovirus type 5; Adenovirus type 7 (7A); Enterovirus Group A (71)(2003); Enterovirus group D (68); Epstein-Barr virus (B95-8); Measles virus; Human cytomegalovirus; Rotavirus, WA strain; Mumps virus 1; Varicella-zoster virus (strain 82); Metapneumovirus (Peru6-2003); Mycoplasma pneumoniae (M129); Chlamydia pneumoniae (ATCC: VR-1435™); Haemophilus influenzae (ATCC: 49144™); Legionella (ATCC: 33152™); Mycobacterium tuberculosis (ATCC: 25177™); Streptococcus pyogenes (ATCC: 19615™); Streptococcus pneumoniae (ATCC:49619™);

Staphylococcus epidermidis(PCI 1200, ATCC: 12228™); Staphylococcus aureus (ATCC: 12600™); Bordetella pertussis type 5(ATCC: 9340-FZ™); Pneumocystis (W303-Pji strain); Candida albicans (ATCC: 44373) cross-reacted with Artron COVID-19 Antigen Test when the virus content>10⁶PFU/mL and the bacterial content>10⁶CFU/mL, nor did they interfere with the test results. The negative matrix prepared from pooled human nasal wash- representative of normal respiratory microbial flora and 20 negative nasopharyngeal swab specimens from healthy volunteers were detected negative, indicating Artron COVID-19 Antigen Test has good analytical specificity.

• Endogenous/Exogenous Interference Study

There was no interference for potential interfering substances listed below: Mucin (0.5% W/V), Whole blood (4%W/V), Beclomethasone (0.5mg/mL), Dexamethasone (1mg/mL), Flunisolide (5mg/mL), Triamcinolone acetonide (1mg/mL), Budesonide (2mg/mL), Mometasone (2mg/mL), Fluticasone (5%V/V), Naso GEL (NeiMed) (5%V/V), Phenylephrine (10%V/V), Oxymetazoline (10%V/V), Sodium chloride (with preservatives) (10% V/V), Menthol (1.5mg/mL), Benzocaine (1.5 mg/mL), CVS Nasal Spray (Cromolyn) (15%V/V), Zicam (5%V/V), Homeopathic (Alkalol) (1:10), Sore Throat Phenol Spray(15%V/V), alpha interferon (200,000IU/mL), Zanamivir (1mg/mL), Ribavirin (2mg/mL), Oseltamivir (5 mg/mL), Peramivir (2mg/mL), Lopinavir (2mg/mL), Ritonavir (2mg/mL), Abidor (4mg/mL), Levofloxacin (5mg/mL), Azithromycin (1mg/mL), Ceftriaxone (1mg/mL), Meropenem (2mg/mL), Mupirocin (10mg/mL), Tobramycin (4µg/mL), Histamine Dihydrochloride (10mg/mL), Biotin (1mg/mL).

• HOOK Effect

There was no hook effect at 9.55×10⁶ TCID₅₀/mL of SARS-CoV-2 strain USA-WA1/2020.

• Clinical Performance

----- nasopharyngeal swab specimens

A total of 812 nasopharyngeal swab specimens including 108 RT-PCR confirmed SARS-CoV-2 positive and 704 RT-PCR confirmed SARS-CoV-2 negative were sequentially enrolled and tested blindly from Nov 27, 2020-Apr 19, 2021. All the 108 RT-PCR positive specimens were collected from symptomatic patients with 75 patients from 0-3 days post onset of symptoms, 26 patients from 4-7 days post onset of symptoms and 7 patients from >7 days post onset of symptoms. Out of 108 positive symptomatic samples, Artron COVID-19 Antigen Test identified 105 positive cases. The diagnostic sensitivity of symptomatic patients was 97.22%(95%CI: 92.10 -99.42), the diagnostic specificity was 99.72% (95% CI: 98.98- 99.97). Overall agreement is 99.39%(98.57-99.80), the Positive Predictive Value for the symptomatic patients was 98.13% (92.93-99.53) whereas the Negative Predictive Value (NPV) was 99.58% (98.72- 99.86).

Clinical studies in asymptomatic patients using serial testing are ongoing to establish clinical performance. The performance of this test has not yet been clinically validated for use in patients without signs and symptoms of respiratory infection or for serial screening applications. Note that performance may differ in these populations.

The performance of Artron COVID-19 Antigen Test against the comparator RT-PCR reagents

Artron COVID-19 Antigen Test	RT-PCR		Total
	Positive	Negative	
Positive	105	2	107
Negative	3	702	705
Total	108	704	812
Performance with 95% CI	Sensitivity	Specificity	Overall Agreement
	97.22% (92.10 -99.42)	99.72% (98.98 -99.97)	99.39% (98.57-99.80)

Summary of positive rate related to Ct Value

Original Ct value for N gene	Artron COVID-19 Antigen Test: Test Positivity Rate with 95%CI	Original Ct value for N gene	Artron COVID-19 Antigen Test: Test Positivity Rate with 95% CI
<27	73/73 (100%)	<30	93/93(100%)
≥27, <30	20/20 (100%)		(96.11-100.00)

≥30, <32	8/9 (88.89%)	≥30	12/15(80.00%)
≥32	4/6 (66.67%)		(51.91-95.67)

Summary of the positive rate related to days post onset

Days post onset of symptoms	Number of Cases	Artron COVID-19 Antigen Test: Test Positivity Rate with 95%CI
0-3	75	75/75(100%) (95.20-100.00)
4-7	26	25/26(96.15%) (80.36-99.90)
>7	7	5/7(71.43%) (29.04-96.33)

-----Nasal swab specimens

A total of 296 cases were recruited into clinical evaluation of Artron COVID-19 Antigen Test with nasal swab specimens. The participants were sequentially enrolled and tested blindly from Jan. 25, 2021 to Mar. 15, 2021. All the 296 cases were confirmed with SARS-CoV-2 RT-PCR at the same timepoint, including 69 SARS-CoV-2 symptomatic positives and 227 SARS-CoV-2 negatives.

Among the 69 positive symptomatic cases, there were 32 patients with 0-3 days post onset of symptoms, 31 patients with 4-7 days post onset, and 6 patients with post onset more than 7 days.

Out of a total of 69 RT-PCR confirmed symptomatic positive cases, Artron COVID-19 Antigen Test was able to correctly detect 63 specimens, with a sensitivity of 91.30% (95%CI: 82.03-96.74); detected 6 cases out of 10 specimens having Ct value over 30, with a sensitivity of 60.00% (6/10, 95%CI: 26.24-87.84); detected 57 specimens from 59 cases with a Ct value below 30 with 96.61% (57/59, 95%CI: 88.29-99.59) sensitivity. Among the 69 positive symptomatic cases, Artron COVID-19 Antigen Test identified 32 from 32 samples with post onset 0-3 days (32/32, 100%, 95%CI: 89.11-100.00), 28 from 31 samples with post onset 4-7 days (28/31, 90.32%, 95%CI: 74.25-97.96), 3 from 6 samples with post onset over 7 days (3/6, 50%, 95%CI: 11.81-88.19).

Artron COVID-19 Antigen Test was able to detect 226 negatives from 227 RT-PCR confirmed negative cases accurately, with a specificity of 99.56% (95%CI: 97.57-99.99).

The Positive Predictive Value (PPV) for all the samples was 98.44% (95%CI: 89.90- 99.78) whereas the Negative Predictive Value (NPV) was 96.97% (95%CI: 94.50-98.54). The overall agreement was 96.97% (95%CI: 94.50-98.54).

The performance of Artron COVID-19 Antigen Test against the comparator RT-PCR reagents

Artron COVID-19 Antigen Test	RT-PCR		Total
	Positive	Negative	
Positive	63	1	64
Negative	6	226	232
Total	69	227	296
Performance with 95% CI	Sensitivity	Specificity	Overall Agreement
	91.30% (82.03-96.74)	99.56% (97.57-99.99)	96.97% (94.50-98.54)

Summary of positive rate related to Ct Value

Original Ct value	Artron COVID-19 Antigen Test: Test Positivity Rate with 95%CI	Original Ct	Artron COVID-19 Antigen Test: Test Positivity Rate with 95% CI
<27	44/44(100%)	<30	57/59(96.61%) (88.29-99.59)
≥27, <30	13/15 (86.67%)		
≥30, <32	5/5 (100%)	≥30	6/10(60%) (26.24-87.84)
≥32	1/5 (0%)		

Summary of the positive rate related to days post onset

Days post onset of symptoms	Number of Cases	Artron COVID-19 Antigen Test: Test Positivity Rate with 95%CI
0-3	32	32/32(100%) (89.11-100.00)
4-7	31	28/31(90.32%) (74.25-97.96)
>7	6	3/6(50%) (11.81-88.19)

REFERENCES

- Clinical management of severe acute respiratory infection (SARI) when COVID-19 disease is suspected. Interim guidance. World Health Organization. 13 March 2020.
- Report of the WHO-China Joint Mission on Coronavirus Disease 2019 (COVID-19). World Health Organization. 16-24 February 2020.
- The Epidemiological Characteristics of an Outbreak of 2019 Novel Coronavirus Diseases (COVID-19). Chinese Center for Disease Control and Prevention. CCDC Weekly, 2(8):113-122, 2020.
- A novel coronavirus outbreak of global health concern. Wang C et al. Lancet, 395(10223):470- 473, 2020

INDEX OF SYMBOLS



Do not reuse



In vitro diagnostic medical device



Temperature limitation



Caution



Manufacturer



Authorised representative in the European community



Batch code



Use by



Contains sufficient for < n > tests



Catalog number



Consult instructions for use



CE Mark

MANUFACTURER CONTACT INFORMATION



Artron

Artron Laboratories Inc.

3938 North Fraser Way

Burnaby, BC

V5J 5H6 Canada

Ph: +1 604-415-9757

Fax: +1 604-415-9795

www.artronlab.com

info@artronlab.com



MedNet EC-REP

GmbH Borkstrasse 10

48163 Muenster

Germany